



**STANDARD OPERATING
PROCEDURE
APTAMARKER KIT**

TABLE OF CONTENTS

Table of Contents	1
Kit Information	2
Kit Contents and Storage	2
Equipments and Reagents Required but not Supplied	2
Regular Laboratory Equipment.....	2
Equipment Specific to Aptamarker Kit	3
Part A : Aptamarker Processing	3
A.1. Preparation of the Aptamarker Library	4
A.2. Application of Sample	4
A.3. Separation of Bound and Unbound Aptamarkers	4
Part B : NGS Preparation	5
B.1. NGS Primer Plate Preparation	5
B.2. Amplification of Libraries.....	6
B.3. Amplified Library Pooling and Cleanup	7
Part C : NGS Submission Recommendations	8
C.1. Determining Number of Useable Reads.....	8
C.2. Sequencing Run Parameters.....	9

KIT INFORMATION

The Aptamarker kit is shipped at room temperature. Upon arrival, all components should be stored at 2-8°C to guarantee maximal shelf life. The kit remains stable at 2-8 °C until the expiration date if stored as specified.

Important Note: Product expires 3 months from the date of shipment.

KIT CONTENTS AND STORAGE

The Aptamarker kit contains the following:

- Aptamarker library plate
- Competitor antisense conjugated to resin solution (Antisense)
Note: 96 samples/tube

- NGS preparation plate
Note: Primers for NGS amplification

Important Note: Store the Aptamarker library plate and the NGS preparation plate at 4°C for up to three months. Store the competitor antisense resin at 4°C for up to two months.

EQUIPMENTS AND REAGENTS REQUIRED BUT NOT SUPPLIED

REGULAR LABORATORY EQUIPMENT

- 1.5 mL microcentrifuge sterile tubes
- Filter tips to minimize potential for contamination
- 5 mL tube
- Nuclease-free water
- Selection buffer PBS 10X
- PCR plate centrifuge
- PCR Plate Sealing Film, adhesive, optical
- Real-Time qPCR Detection System for characterization of amplification
- Two separate laboratory benches

EQUIPMENT SPECIFIC TO APTAMARKER KIT

- One 96 well plate (200 µL minimum volume)
- Cleanup column : GFX™ PCR DNA and Gel Band Purification Kit : 28903466, 28903470 or 28903471 (Cytiva) One cleanup column per NGS set submission.
- qPCR Master Mix (SsoAdvanced Universal SYBR® Green Supermix : 1725270 to 1725275 (BioRad)

PART A : APTAMARKER PROCESSING

- Procedures within this section are performed at room temperature (RT). Aptamarker library plate may remain open during processing and should be sealed for centrifuging.
- The Aptamarker library plate contains an average of 10,000 copies of each of 268M different Aptamarker probes in each well. Given the need for 20M reads per sample for next generation sequencing analysis, the plate should be considered as two separate sets for Aptamarker processing. Appropriate labelling of the plate is provided in Figure 1.

Figure 1: Aptamarker library plate processing guide.

	1	2	3	4	5	6	7	8	9	10	11	12
A	1	9	17	25	33	41	49	57	65	73	81	89
B	2	10	18	26	34	42	50	58	66	74	82	90
C	3	11	19	27	35	43	51	59	67	75	83	91
D	4	12	20	28	36	44	52	60	68	76	84	92
E	5	13	21	29	37	45	53	61	69	77	85	93
F	6	14	22	30	38	46	54	62	70	78	86	C1
G	7	15	23	31	39	47	55	63	71	79	87	C2
H	8	16	24	32	40	48	56	64	72	80	88	C3

Example of Sample/well pairings for 93 samples and 3 control libraries.

Numbers refer to Sample number and C1 to C3 refer to library controls.

A.1. PREPARATION OF THE APTAMARKER LIBRARY

1. Briefly centrifuge the plate and unseal it.
2. Add 20 μL of nuclease-free water and 5 μL of 10X selection buffer to each used well of Aptamarker library plate. Pipette up and down once to mix the solutions.
3. Incubate for 30 min at RT.

A.2. APPLICATION OF SAMPLE

1. Add 10 μL of sample (for control libraries, sample is replaced with water) to the 25 μL of prepared Aptamarker library and buffer by pipetting up and down to mix the solution.
2. Incubate for 30 minutes at RT.
3. Add 15 μL of competitor Antisense resin to each tube containing 35 μL of Aptamarker library, sample and buffer solution.

Important Note: The stock tube should be mixed and vortexed to resuspend resin prior to use.

4. Seal the plate with fresh film.
5. Incubate for 15 minutes at RT agitating the plate manually every five minutes to prevent the Antisense resin from settling.

A.3. SEPARATION OF BOUND AND UNBOUND APTAMARKERS

1. Centrifuge the plate at 10 000 x g for 10 minutes to pellet Aptamarkers that are hybridized to Antisense on resin.
2. Unseal the plate carefully.
3. Transfer 30 μL of the supernatant to a clean 96-well plate for dilution (this plate is not provided in the kit).

Important Note: It is important not to disturb the resin at the bottom of the tube while collecting the supernatant.

4. Add 120 μL of Nuclease-free water to the collected supernatant and pipette up and down once to mix the solutions.
5. Discard the Aptamarker library plate.
6. Either seal the dilution plate and store the plate at 4°C or proceed directly to NGS preparation.

PART B : NGS PREPARATION

- The NGS preparation plate has different forward primers in each 96 wells. All wells contain the same reverse primer. A guide for the use of this plate is provided in Figure 2.
- **Maintain alignment of samples processed in the Aptamarker processing plate with the NGS processing plate.**
- The amplified product from each 96 wells is pooled for one NGS submission lane.
- Part B.1. should be performed on another different clean bench than part A and B.2. and B.3..

Figure 2: Outline of project plan for 93 samples and 3 control libraries per NGS submission lane:

	1	2	3	4	5	6	7	8	9	10	11	12
A	1	9	17	25	33	41	49	57	65	73	81	89
B	2	10	18	26	34	42	50	58	66	74	82	90
C	3	11	19	27	35	43	51	59	67	75	83	91
D	4	12	20	28	36	44	52	60	68	76	84	92
E	5	13	21	29	37	45	53	61	69	77	85	93
F	6	14	22	30	38	46	54	62	70	78	86	C1
G	7	15	23	31	39	47	55	63	71	79	87	C2
H	8	16	24	32	40	48	56	64	72	80	88	C3

Sample/well pairings for 93 samples and 3 control libraries.

Numbers refer to Sample number and C1 to C3 refers to library controls.

B.1. NGS PRIMER PLATE PREPARATION

1. On a clean separate bench, briefly centrifuge the plate and unseal it.
2. Add 10 µL of nuclease-free water to each used well within the NGS preparation plate.
3. Incubate for 30 min at RT.
4. Pipette 20 µL of SYBR Green master mix into each well of the NGS processing plate

Important Note: To ensure optimal performance, make sure SYBR Green master mix is thawed and homogenized before use.

B.2. AMPLIFICATION OF LIBRARIES


1. Pipette 10 μ L of template (supernatant diluted in water obtained in Part A) into each respective well to a final volume of 40 μ L.

Important Note: To avoid cross-contamination, we strongly recommend pipetting the template last, preferably in a separate work area.

2. Seal the plate with a fresh qPCR appropriate optical adhesive film before proceeding with the real-time qPCR detection steps.
3. Centrifuge the plate to spin down the contents and eliminate air bubbles.
4. Load the reaction plate onto the real-time qPCR instrument and start the qPCR run with Sybr Green setting/mode for a final volume of 40 μ L with the protocol defined below:

Figure 3: qPCR protocol

STEP	TEMPERATURE	TIME
Polymerase Activation and Initial Denaturation	95°C	5 minutes
<u>4 Cycles</u>	95°C	10 seconds
	<u>50°C</u>	15 seconds
	72°C	30 seconds
<u>16 Cycles</u>	95°C	10 seconds
	<u>64°C</u>	15 seconds
	72°C + Plate Read	30 seconds
Final Extension	72°C	5 minutes
Hold	4-10°C	Infinite



Important Note: Once the amplification step is performed, keep plate stored at 4°C and keep plate away from the light.

B.3. AMPLIFIED LIBRARY POOLING AND CLEANUP

1. Pipette 10 µL of each amplified library within the 96 wells into one clear 5 mL tube.
2. Proceed to cleanup by following the protocol of the Cytiva for the purification of DNA from solution or an enzymatic reaction from the supplier.

Important Note: Use a separate Cytiva cleanup column for each amplified library pool.

Important Note: Cytiva binding buffer volume varies depending on volume of library pool.

Important Note: We have only qualified the Cytiva columns for this application given the relative length of the primers.

3. After the full amplified library pool has been passed through a cleanup column elute the libraries with 150 µL of Nuclease-free water.

PART C : NGS SUBMISSION RECOMMENDATIONS

C.1. DETERMINING NUMBER OF USEABLE READS

We estimate that 64% of the raw instrument output will result in high-quality usable reads for downstream analysis. The Aptamarker library has low genetic diversity, so we must account for data loss required to maintain high sequencing quality. With the low diversity library, we recommend that it be spiked with ~15% PhiX to provide the necessary sequence complexity for NovaSeq sensors.

Factors impacting total useable reads:

Factor	Value	Rationale
Q30 Quality	90%	A 10% loss is budgeted to ensure we only analyze sequences with base calls with 99.9% accuracy.
PhiX	85%	Essential for low-diversity libraries. Spike with 15% PhiX, leaving 85% of the flow cells for your samples.
Matching	86%	Accounts for reads that are not perfectly aligned with sequences in the Aptamarker library.
Deconvoluting	97%	The expected efficiency of the demultiplexing process (assigning reads to their specific hex codes).
Expected useable reads	64%	Total calculation: $0.90 \times 0.85 \times 0.86 \times 0.97 \approx 64\%$

Useable reads are calculated as follows:

$$\text{Reads per sample}_{\text{Expected}} = \frac{(\text{Total Flow Cell Capacity} \times 0.64)}{\text{Total Libraries}}$$

Description:

1. Multiply the expected number of total reads for your NGS instrument by the expected number of useable reads (0.64).
2. Divide by the number of libraries included in each NGS submission (samples and technical controls).

This will provide you with the expected number of reads/samples. A minimum of 20 million (M) useable reads per sample is required. This depth is necessary to capture the full complexity of the Aptamarker library and provide statistically significant results during downstream analysis.

C.2. SEQUENCING RUN PARAMETERS

To maximize read depth and efficiency, the run must be configured as follows:

- **Mode:** One-way (Single-End) sequencing
- **Cycles:** 100 cycles

The Aptamarker library was designed and validated around using single-end sequencing. To obtain 20M reads/sample, it is essential to run NGS using the recommended mode.

Users intending to use NovaSeq X 25B flow cells should contact us and order an Aptamarker kit designed for this type of analysis. Kits for NGS analysis with NovaSeq X 25B flow cells contain an NGS preparation plate with 96 different forward primers allowing analysis of 93 samples in one NGS submission.

This product is intended for research purposes only. This product is not intended for therapeutic or diagnostic purposes in humans or animals.

This product is covered by one or more patents, trademarks and/or copyrights owned by NeoVentures Biotechnology Inc. and NeoVentures Biotechnology Europe SAS.

NeoVentures Biotechnology Europe SAS – 1, Mail du Professeur Georges Mathé – 94800 Villejuif – +33 (0)6 86 98 36 95

<https://neovertures-eu.com/>

Service & support requests : slecocq@neoneuro-aptamers.com